

HOE VERTEL JE DIT AAN JE CONSUMENTEN?

1

VANAF 40 JAAR, VERANDERT ONS ZICHT. DIT NOEMEN WE PRESBYOPIE.

Om een goed zicht te behouden en voluit deel te nemen aan dagelijkse activiteiten, zijn gepaste oplossingen of corrigerende maatregelen essentieel.

2

VERSCHILLENDE OPTIES OM ZICHT TE CORRIGEREN VANAF 40 JAAR



- ▶ **Leesbril**
U moet hem telkens op- en afzetten als u nabij of wel veraf scherp wil zien.
- ▶ **Multifocale bril**
Hij is ontworpen om zowel nabij, tussenin als veraf scherp te zien, maar u moet hem wel voortdurend dragen.
- ▶ **Contactlenzen voor presbyopie**
U hoeft geen bril te dragen, want deze contactlens is ontworpen om zowel nabij, tussenin als veraf scherp zicht te bieden.

3

WAAROM KIEZEN VOOR CONTACTLENZEN?



In tegenstelling tot een leesbril bieden multifocale contactlenzen zowel helder zicht¹⁻³ als comfort^{4,11} in verschillende alledaagse situaties, zoals tijdens het sporten of voor een avondje uit.



TOTAL® Multifocal is specifiek ontworpen om presbyopie te corrigeren. Deze contactlenzen bieden **helder zicht** op alle afstanden, van dichtbij tot veraf¹⁻³ en **uitzonderlijk comfort**^{4,11} dankzij de unieke Water Surface-technologie met bijna 100% water aan het lensoppervlak.^{7,10*}



Met DAILIES TOTAL1® Multifocal daglenzen kunt u kiezen op basis van uw levensstijl of activiteiten, zonder in te leveren op zicht of comfort.²⁻⁵ Ze zijn perfect als u houdt van dagelijks een **nieuw paar contactlenzen**. Ze zijn ook ideaal voor alledaags of occasioneel gebruik, of zelfs om te sporten, te werken en om uit te gaan. TOTAL30® Multifocal contactlenzen zijn perfect als u liever **hetzelfde paar contactlenzen draagt gedurende 30 dagen**.

MULTIFOCAL CONTACTLENZEN VOOR ULTIEM COMFORT^{6,11}



- **Alcon's nieuwste** generatie contactlenzen^{7,9,12}
- **-100% water aan het oppervlak**^{7,10*} waardoor enkel een zacht kussentje van vocht het oog raakt.
- **Het lensoppervlak blijft langer bevochtigd**^{13-18**†}
- **5 keer zachter** dan andere contactlenzen^{19-21††,‡}

- * Gebaseerd op in vitro metingen van ongedragen lenzen.
- ** Based on an in vitro study on TOTAL30®, Biofinity, Ultra, ACUVUE OASYS and ACUVUE VITA sphere lenses wherein wettability was measured using the iDDrop System (p<0.05). All lenses were tested in an identical manner, soaked in a PBS (phosphate-buffered saline solution) for 16 hours plus/minus 2 hours (p<0.05).
- † Based on in vitro studies on DAILIES TOTAL1®, INFUSE, ACUVUE OASYS 1-DAY, ACUVUE OASYS MAX 1-DAY, BioTrue ONEday, 1-DAY ACUVUE MOIST, clariti 1 day and MyDay contact lenses wherein wettability was measured using the iDDrop System (p<0.001). All lenses were tested in an identical manner, soaked in a PBS (phosphate-buffered saline solution) for 16 hours +/- 2 hours (p<0.001).
- †† Based on surface modulus measured with AFM nanoindentation ("worn") sphere lenses were worn on-eye for a full wearing schedule: 14 days for ACUVUE OASYS®, 30 days for Biofinity®, ACUVUE VITA®, Bausch+Lomb ULTRA®; p<0.05.
- ‡ Based on surface modulus measured with AFM nanoindentation studies with delefilcon A material (DAILIES TOTAL1 sphere lenses); compared to ACUVUE OASYS 1-DAY, ACUVUE OASYS MAX 1-DAY, 1-DAY ACUVUE MOIST, clariti 1 day and MyDay sphere contact lenses; p<0.01.

- Referenties:
1. Based on a clinical performance evaluation of TOTAL30® Multifocal contact lenses; Alcon data on file, 2022.
 2. Lemp J, Kern J. Alcon multifocal contact lenses for presbyopia correction. Paper presented at the Canadian Association of Optometrists Congress; June 28-30, 2017; Ottawa, ON.
 3. Merchea M, Evans D, Kannarr S, et al. Patient and ECP Satisfaction in the United States with a Novel Water Gradient Daily Disposable Multifocal Contact Lens. Presented at the American Optometric Association Annual Meeting, June 20-24, 2018, Denver, CO.
 4. Based on an assessment of DAILIES TOTAL1® Multifocal contact lenses in challenging patients; Alcon data on file, 2017.
 5. Based on a clinical trial of DAILIES TOTAL1® Multifocal contact lenses; Alcon data on file, 2013.
 6. In a clinical study wherein patients used AOSEPT® solution for nightly cleaning, disinfecting, and storing; Alcon data on file, 2021.
 7. In vitro analysis of lens oxygen permeability, water content, and surface imaging; Alcon data on file, 2021.
 8. In vitro analysis of delefilcon A contact lenses outermost surface softness and correlation with water content; Alcon data on file, 2020.
 9. Thekveli S, Qui Y, Kapoor Y, et al. Structure-property relationship of delefilcon A lenses. Contact Lens Anterior Eye. 2012;35(Suppl 1):e14.
 10. Angelini TE, Nixon RM, Dunn AC, et al. Viscoelasticity and mesh-size at the surface of hydrogels characterized with microrheology. Invest Ophthalmol Vis Sci. 2013;54:E-abstract 500.
 11. Perez-Gomez I, Giles T. European survey of contact lens wearers and eye care professionals on satisfaction with a new water gradient daily disposable contact lens. Clin Optom. 2014;6:17-23.
 12. Shi X, Cantu-Crouch D, Sharma V, et al. Surface characterization of a silicone hydrogel contact lens having bioinspired 2-methacryloyloxyethyl phosphorylcholine polymer layer in hydrated state. Colloids Surf B: Biointerfaces. March 2021;199:111539.
 13. Based on in vitro studies of unworn TOTAL30® and competitive sphere lenses (Biofinity, Ultra, ACUVUE OASYS and ACUVUE VITA) using average iDDrop water break-up time (WBUT); Alcon data on file, 2021.
 14. Based on an in vitro study on unworn DAILIES TOTAL1® and competitive sphere lenses (1-DAY ACUVUE MOIST, MyDay and clariti 1 day) using average iDDrop water break-up time (WBUT); Alcon data on file, 2019.
 15. Based on an in vitro study on unworn ACUVUE OASYS 1-DAY and ACUVUE TrueEye sphere lenses using average iDDrop water break-up time (WBUT); Alcon data on file, 2019.
 16. Based on an in vitro study on unworn infuse sphere lenses testing average iDDrop water break-up time (WBUT) and surface softness; Alcon data on file, 2020.
 17. Based on an in vitro study on unworn DAILIES® AquaComfort PLUS® sphere and BioTrue ONEday sphere lenses using average iDDrop water break-up time (WBUT); Alcon data on file, 2019.
 18. Based on an in vitro study on unworn ACUVUE OASYS MAX 1-DAY sphere lenses using average iDDrop water break-up time (WBUT); Alcon data on file, 2022.
 19. Laboratory analysis of surface modulus of delefilcon A sphere and commercial sphere lenses using atomic force microscope; Alcon data on file, 2020, 2021.
 20. Laboratory analysis of surface modulus of delefilcon A sphere lenses and commercial sphere lenses using atomic force microscope; Alcon data on file 2020.
 21. Laboratory analysis of surface modulus of senofilcon A sphere lenses (ACUVUE OASYS MAX 1-DAY) using atomic force microscope; Alcon data on file, 2022.
 22. Alcon data on file, 2022.
 23. Based on an assessment of DAILIES TOTAL1® Multifocal contact lenses in challenging patients; Alcon data on file, 2017.
 24. Based on Euromcontact Q1 2023 Benelux.
 25. Multi-sponsor Surveys, Inc. 2022 Syndicated Study of the U.S. Consumer Contact Lens Market.
 26. Sulley A, Young G, Hunt C. Factors in the success of new contact lens wearers. Cont Lens Anterior Eye. 2017;40(1):15-24.
 27. 2021 Alcon U.S. consumer presbyopia study.
 28. Bauman E, Lemp J, Kern J. Material effect on multifocal contact lens fitting of lenses of the same optical design with the same fitting guide. Poster presented at: British Contact Lens Association Clinical Conference & Exhibition; June 9-11, 2017; Liverpool, UK.

De aanpassing van contactlenzen in België is voorbehouden aan optometristen/orthoptisten zoals beschreven in het Koninklijk Besluit van 7 oktober 2022. Alcon kan niet verantwoordelijk worden gehouden voor enig ongeoorloofd gebruik. Voor indicaties, contra-indicaties en waarschuwingen verwijzen wij u naar de gebruiksaanwijzing van het betreffende product. Contactlenzen zijn medische hulpmiddelen. © 2023 Alcon Inc. BNLX-VC-2300036 12/2023

MULTIFOCAAL

MEER KLANTEN MET PRESBYOPIE VERDIENEN
HELDER ZICHT¹⁻³
EN EEN ULTIEME LENSERVARING³⁻⁶



Alcon

Alcon

DE MARKT VAN MULTIFOCALE LENZEN



Multifocaal is het **snelst groeiende segment** in de maandlenssector in de Benelux.²⁴



Multifocale lenzen zijn essentieel voor **klantretentie na de leeftijd van 40 jaar** omdat het risico op drop-out toeneemt.^{25,26}



Klanten met presbyopie hechten evenveel waarde aan het **comfort** van multifocale contactlenzen als aan **scherp zicht**.²⁵



De meeste klanten wachten meer dan een jaar voordat ze presbyopie met u zullen bespreken.²⁷



De **aarzeling om een leesbril te dragen** kan gevoelens van frustratie, schaamte en hulpeloosheid met zich meebrengen.

62%


van nieuwe klanten met presbyopie is bereid te betalen voor een oplossing die verder gaat dan alleen een parttime bril.²⁷

DE TOTAL™ FAMILY IS EINDELIJK COMPLEET




BEPROEFD LENS MATERIAAL

Het oog komt enkel in contact met een kussen van vocht.⁷⁻¹⁰
Het resultaat is **ULTIEM COMFORT**^{6,11}




ClassicSiHy lenses

Homogenous lens material¹. No water surfacing technology*



DAILIES TOTAL1™ daily lenses

~100% water content on the surface^{9,10*}



TOTAL30™ monthly lenses

~100% water content on the surface^{7,8*}



Dichtbij
Tussenliggende afstanden
Veraf


BEPROEFD PRECISION PROFILE™-LENSONTWERP

Uitstekend zicht van dichtbij over tussenliggende afstanden tot veraf^{1,2}



BEWEZEN AANPASPROCES^{3,22,28}

Hetzelfde aanpasproces voor alle **TOTAL™ multifocale lenzen**



195
parameters

Sferische sterktes
+6.00 tot -10.00D
(stappen van 0.25D)

ADD-sterktes
<1.25D (LOW)
1.50D tot 2.00D (MED)
>2.25 (HIGH)



96%
SUCCES BIJ AANPASSING

SUCCES BIJ AANPASSING MET 2 LENZEN OF MINDER PER OOG^{3,22,28*}

*Met twee of minder lenzen bij het eerste pasbezoek

BENUT DE VELE KANSSEN BIJ UW KLANTEN

SWITCH

dragers van AIR OPTIX™ plus HydraGlyde™ Multifocal- en DAILIES™ AquaComfort™ PLUS Multifocal-contactlenzen



LAAT LENS DRAGERS OVERSTAPPEN

van sferische naar multifocale contactlenzen



LAAT NIEUWE BRILDRAGERS KENNISMAKEN

om hun bril met lenzen af te wisselen

